








SESAA Esterhazy Station - August 2026 Summary Report

Pollutant		Month Records		24 Hour Records				1 Hour Records			
Name	Unit	Average Concentration	Uptime	Maximum		SAAQS*	Exceedance Days	Maximum		SAAQS*	Exceed Hours
				Conc.	Date			Concentration	Date and Time		
NO	ppb	0.6	79.6%	0.8	8/10/2026	-	-	5.3	08/27/2026 08:00	-	-
NO ₂	ppb	0.5	79.6%	0.9	8/11/2026	106	0	5.9	08/18/2026 21:00	159	0
NO _x	ppb	1.1	79.6%	1.7	8/11/2026	-	-	6.4	08/18/2026 21:00	-	-
O ₃	ppb	19.4	91.7%	38.8	8/24/2026	63 (8-Hrs Rolling Avg)	0	43.3	08/24/2026 14:00	82	0
PM _{2.5}	µg/m ³	6.8	27.0%	10.2	8/29/2026	28	0	32.4	08/07/2026 13:00	-	-

SAAQS: Saskatchewan Ambient Air Quality Standards




*If AQHI cannot be calculated due to missing parameters, AQHI plus is reported using PM_{2.5}




Exceedance Summary (24-Hour Average)										
			Summary of other Parameters during Exceedance							
Pollutant	Date	Conc.	WS	WD	Precip.	NO	NO ₂	NO _x	O ₃	PM _{2.5}

Exceedance Summary (1-hour Average)										
			Summary of other Parameters during Exceedance							
Pollutant	Date	Conc.	WS	WD	Precip.	NO	NO ₂	NO _x	O ₃	PM _{2.5}


SESAA Esterhazy Station - August 2026 Summary Report




Wind Rose



Pollution Rose NO₂



Pollution Rose O₃



Pollution Rose PM_{2.5}

Parameters	Unit	Color Code for Wind Speed and Pollutant Concentration Class						
Wind Speed	m/s	0.5 - 1.2	1.2 - 2.4	2.4 - 3.6	3.6 - 4.8	4.8 - 6.0	C> 6.0	
NO ₂	ppb	0.0-C-4.0	4.0-C-8.0	8.0-C-12.0	12.0-C-16.0	16.0-C-20.0	C> 20.0	
O ₃	ppb	0.0-C-10.0	10.0-C-20.0	20.0-C-30.0	30.0-C-40.0	40.0-C-50.0	C> 50.0	
PM _{2.5}	µg/m ³	0.0-C-10.0	10.0-C-20.0	20.0-C-30.0	30.0-C-40.0	40.0-C-50.0	C> 50.0	

Monthly Update










- * All data has been zero drift, and baseline corrected, when required. Data may change after final validation process.
- * All compliance parameters were within the Saskatchewan Ambient Air Quality Standards.
- * All compliance parameters were less than 90% operational for the month with the exception of O₃.
- * A total of 62 hours of O₃ data was unavailable due to power issue.
- * A total of 543 hours of PM_{2.5} data was unavailable due to analyzer error and power issue.
- * A total of 152 hours of NO₂ data was unavailable due to analyzer error as well as power issue.
- * Quarterly calibrations completed on June 20, 2026



SESAA Glen Ewen Station - August 2026 Summary Report

Pollutants		Month Records		24 Hour Records				1 Hour Records			
Name	Unit	Average Concentration	Uptime	Maximum		SAAQS*	Exceedance Days	Maximum		SAAQS*	Exceed Hours
				Conc.	Date			Conc.	Date and Time		
NO	ppb	1.8	75.7%	1.9	8/10/2026	-	-	10	08/27/2026 07:00	-	-
NO ₂	ppb	0.9	75.7%	1.4	8/10/2026	106	0	5	08/27/2026 07:00	159	0
NO _x	ppb	2.7	75.7%	3.3	8/10/2026	-	-	15	08/27/2026 07:00	-	-
SO ₂	ppb	0.2	95.2%	0.5	8/2/2026	48	0	4.2	08/06/2026 11:00	172	0
H ₂ S	ppb	2.1	95.2%	3.0	8/8/2026	3.6	0	11.8	08/08/2026 05:00	11	1
O ₃	ppb	22.2	95.4%	54.7	8/1/2026	63 (8 hrs rolling avg)	0	56	08/01/2026 16:00	82	0

SAAQS: Saskatchewan Ambient Air Quality Standards








SESAA Glen Ewen Station - August 2026 Summary Report

Exceedance Summary (24-Hour Average)

24-hour Exceedance Information			Summary of other Parameters during Exceedance							
Pollutant	Date	Conc.	WS	WD	NO	NO ₂	NO _x	SO ₂	H ₂ S	O ₃

Exceedance Summary (1-Hour Average)

1-hour Exceedance Information			Summary of Other Parameters During Exceedance					
Pollutant	Date and Time	Conc.	WS	WD	Precip.	SO ₂	NO ₂	O ₃
H ₂ S	8/8/26 5:00 AM	11.8	0.63	266	0.0	0.0	1.5	5.7



Parameters	Unit	Color Code for Wind Speed and Pollutant Conc. Class						
Wind Speed	m/s	0.5-W- 1.2	1.2 -W- 2.4	2.4 -W- 3.6	3.6 -W- 4.8	4.8 -W- 6.0	W > 6.0	
NO ₂	ppb	0.0-C- 1.2	1.2 -C- 2.4	2.4 -C- 3.6	3.6 -C- 4.8	4.8 -C- 6.0	C > 6.0	
SO ₂	ppb	0.0-C- 1.2	1.2 -C- 2.4	2.4 -C- 3.6	3.6 -C- 4.8	4.8 -C- 6.0	C > 6.0	
H ₂ S	ppb	0.0-C- 1.2	1.2 -C- 2.4	2.4 -C- 3.6	3.6 -C- 4.8	4.8 -C- 6.0	C > 6.0	
O ₃	ppb	0.0-C-8.0	8.0-C-16.0	16.0-C-24.0	24.0-C-32.0	32.0-C-40.0	C > 40.0	

Monthly Update


- * All data has been zero drift and baseline corrected, when required. Data may change after final validation process.
- * All compliance parameters were within the Saskatchewan Ambient Air Quality Standards with the exception of one 1-hour H₂S exceedance.
- * All compliance parameters were more than 90% operational for the month with the exception of NO_x.
- * A total of 181 hours of NO₂ data unavailable due to analyzer error.
- * One hour of data were flagged for all gas analyzers due to daily IZS checks, resulting in the loss of one hour of data each day.
- * Shelter temperature instability during the reporting period affected analyzers internal checks and performance, resulting in intermittent data loss and compromising quality of data.
- * Quarterly calibrations were completed on June 18, 2026.











SESAA Oxbow Station - August 2026 Summary Report

Pollutants		Month Records		24 Hour Records				1 Hour Records			
Name	Unit	Average Concentration	Uptime	Maximum		SAAQS*	Exceedance Days	Maximum		SAAQS*	Exceed Hours
				Conc.	Date			Conc.	Date and Time		
NO	ppb	0	95.6%	0.2	8/26/2026	-	-	2.3	08/26/2026 08:00	-	-
NO ₂	ppb	4.2	95.6%	5.2	8/31/2026	106	0	11.5	08/12/2026 23:00	159	0
NO _x	ppb	3.9	95.6%	5.1	8/11/2026	-	-	13.4	08/11/2026 07:00	-	-
SO ₂	ppb	0.4	95.6%	0.9	8/11/2026	48	0	6.2	08/08/2026 11:00	172	0
H ₂ S	ppb	1.2	95.6%	1.7	8/3/2026	3.6	0	3.6	08/03/2026 09:00	11	0
PM _{2.5}	µg/m ³	14.1	88.2%	32.2	8/5/2026	28	1	65	08/05/2026 08:00	-	-

SAAQS: Saskatchewan Ambient Air Quality Standards




*If AQHI cannot be calculated due to missing parameters, AQHI plus is reported using PM_{2.5}



SESAA Oxbow Station - August 2026 Summary Report

Exceedance Summary

Exceedance Information			Summary of other Parameters during Exceedance							
Pollutant	Date	Conc.	WS	WD	H ₂ S	NO	NO ₂	NO _x	SO ₂	PM _{2.5}
PM _{2.5}	5-Aug	32.2	3.9	230.6	1.6	0.0	5.2	4.9	0.7	32.2



Parameters	Unit	Color Code for Wind Speed and Pollutant Concentration Class						
Wind Speed	m/s	0.5 -W- 1.2	1.2 -W- 2.4	2.4 -W- 3.6	3.6 -W- 4.8	4.8 -W- 6.0	W > 6.0	
NO ₂	ppb	0.0-C- 1.2	1.2 -C- 2.4	2.4-C- 3.6	3.6-C- 4.8	4.8 -C- 6.0	C > 6.0	
SO ₂	ppb	0.0-C-2.0	2.0-C-4.0	4.0-C-6.0	6.0-C-8.0	8.0-C-10.0	C > 10.0	
H ₂ S	ppb	0.0-C-2.0	2.0-C-4.0	4.0-C-6.0	6.0-C-8.0	8.0-C-10.0	C > 10.0	
PM _{2.5}	µg/m ³	N/A	N/A	N/A	N/A	N/A	N/A	

Monthly Update


- * All data has been zero drift and baseline corrected, when required. Data may change after final validation process.
- * All compliance parameters were within the Saskatchewan Ambient Air Quality Standards with the exception of one, 24-hour PM_{2.5} exceedance.
- * All compliance parameters were more than 90% operational for the month, except for PM_{2.5}.
- * A total of 88 hours of PM_{2.5} data is missing due to the analyzer being removed for repair, analyzer reinstalled on August 4, 2026 .
- * One hour of data were flagged for all gas analyzers due to daily IZS checks, resulting in the loss of one hour of data each day.
- * Quarterly calibrations were completed on June 17, 2026.











SESAA Stoughton Station - August 2026 Summary Report

Pollutants		Month Records		24 Hour Records				1 Hour Records			
Name	Unit	Average Concentration	Uptime	Maximum		SAAQS*	Exceedance Days	Maximum		SAAQS*	Exceed Hours
				Conc.	Date			Conc.	Date and Time		
NO	ppb	0.1	95.7%	0.5	8/16/2026	-	-	7.5	08/16/2026 18:00	-	-
NO ₂	ppb	0.5	95.7%	1.5	8/11/2026	106	0	9.0	08/11/2026 01:00	159	0
NO _x	ppb	0.4	95.7%	1.7	8/11/2026	-	-	11.4	08/11/2026 01:00	-	-
SO ₂	ppb	0.1	95.7%	0.8	8/8/2026	48	0	8.8	08/08/2026 10:00	172	0
H ₂ S	ppb	0.9	95.7%	1.5	8/22/2026	3.6	0	8.8	08/16/2026 07:00	11	0
PM _{2.5}	µg/m ³	12.1	17.5%	19.5	8/27/2026	28	0	25.7	08/26/2026 21:00	-	-

SAAQS: Saskatchewan Ambient Air Quality Standards



*If AQHI cannot be calculated due to missing parameters, AQHI plus is reported using PM_{2.5}




SESAA Stoughton Station - August 2026 Summary Report

Exceedance Summary (24-hour)


Exceedance Summary			Summary of other Parameters during Exceedance							
Pollutant	Date	Conc.	WS	WD	H ₂ S	NO	NO ₂	NOx	SO ₂	PM _{2.5}

Exceedance Summary (1-hour)


Exceedance Summary			Summary of other Parameters during Exceedance							
Pollutant	Date & Time	Conc.	WS	WD	H ₂ S	NO	NO ₂	NOx	SO ₂	PM _{2.5}




Wind Rose




NO₂ Pollution Rose



SO₂ Pollution Rose



H₂S Pollution Rose



PM_{2.5} Pollution Rose

Parameters	Unit	Color Code for Wind Speed and Pollutant Concentration Class						
Wind Speed	m/s	0.5 -W- 1.2	1.2 -W- 2.4	2.4 -W- 3.6	3.6 -W- 4.8	4.8 -W- 6.0	W > 6.0	
NO ₂	ppb	0.0 -C- 1.2	1.2 -C- 2.4	2.4 -C- 3.6	3.6 -C- 4.8	4.8 -C- 6.0	C > 6.0	
SO ₂	ppb	0.0 -C- 1.2	1.2 -C- 2.4	2.4 -C- 3.6	3.6 -C- 4.8	4.8 -C- 6.0	C > 6.0	
H ₂ S	ppb	0.0 -C- 1.2	1.2 -C- 2.4	2.4 -C- 3.6	3.6 -C- 4.8	4.8 -C- 6.0	C > 6.0	
PM _{2.5}	µg/m ³	0.0-C-3.0	3.0-C-6.0	6.0-C-9.0	9.0-C-12.0	12.0-C-15.0	C > 15.0	

Monthly Update








- * All data has been zero drift and baseline corrected, if needed. Data may change after final validation process.
- * All compliance parameters were within the Saskatchewan Ambient Air Quality Standards.
- * All compliance parameters were over 90% operational for the month, except PM_{2.5}.
- * A total of 614 hours of PM_{2.5} data intermittent throughout the month were flagged as "analyzer malfunction," resulting in missing and flatlined data.
- * One hour of data were flagged for all gas analyzers due to daily IZS checks, resulting in the loss of one hour of data each day.
- * Quarterly calibrations were completed on June 16, 2026.



SESAA Torquay Station - August 2026 Summary Report





Pollutants		Month Records		24 Hour Records				1 Hour Records			
Name	Unit	Average Concentration	Uptime	Maximum		SAAQS*	Exceedance Days	Maximum		SAAQS*	Exceed Hours
				Conc.	Date			Conc.	Date and Time		
SO ₂	ppb	0.2	93.6%	0.7	8/28/2026	48	0	4.2	08/08/2026 12:00	172	0
H ₂ S	ppb	1.5	93.8%	2.0	8/3/2026	3.6	0	10.1	08/04/2026 07:00	11	0
PM _{2.5}	µg/m ³	10.8	96.0%	29.4	8/2/2026	28	1	81.5	08/11/2026 22:00	-	-

SAAQS: Saskatchewan Ambient Air Quality Standards



Exceedance Summary								
24-hour Exceedance Information			Summary of other Parameters during Exceedance					
Pollutant	Date	Conc.	WS	WD	Precip.	SO ₂	H ₂ S	PM _{2.5}
PM _{2.5}	2-Aug	29.4	2.6	262.6	0.0	0.4	1.5	29.4

SESAA Torquay Station - August 2026 Summary Report



Parameters	Unit	Color Code for Wind Speed and Pollutant Concentration Class						
Wind Speed	m/s	0.5 -W- 1.2	1.2 -W- 2.4	2.4 -W- 3.6	3.6 -W- 4.8	4.8 -W- 6.0	W > 6.0	
SO ₂	ppb	0.0-C-0.6	0.6-C-1.2	1.2-C-1.8	1.8-C-2.4	2.4-C-3.0	C > 3.0	
H ₂ S	ppb	0.0-C-0.6	0.6-C-1.2	1.2-C-1.8	1.8-C-2.4	2.4-C-3.0	C > 3.0	
PM _{2.5}	ug/m ³	0.0-C-4.0	4.0-C-8.0	8.0-C-12.0	12.0-C-16.0	16.0-C-20.0	C > 20.0	








Monthly Update

- * All data has been zero drift and baseline corrected, where required. Data may change after final validation process.
- * All compliance parameters were within the Saskatchewan Ambient Air Quality Standards with the exception of one, 24-Hour PM_{2.5} exceedance.
- * All compliance parameters were greater than 90% operational for the month.
- * One hour of data were flagged for all gas analyzers due to daily IZS checks, resulting in the loss of one hour of data each day.
- * Shelter temperature fluctuated during the reporting period and reached unusually high levels, potentially affecting data quality. The affected data were flagged and provided for reference only.
- * Most recent quarterly calibrations were completed on June 19, 2026.



SESAA Wauchope Station - August 2026 Summary Report											
Pollutants		Month Records		24 Hour Records				1 Hour Records			
Name	Unit	Average Concentration	Uptime	Maximum		SAAQS*	Exceedance Days	Maximum		SAAQS*	Exceed Hours
				Conc.	Date			Conc.	Date and Time		
SO ₂	ppb	0.7	73.0%	1.3	8/11/2026	48	0	6	08/11/2026 10:00	172	0
H ₂ S	ppb	2.2	73.7%	5.6	8/11/2026	3.6	3	17.9	08/11/2026 06:00	11	10
PM _{2.5}	µg/m ³	No Data	0.0%	0.0	NA	28	0	No Data	NA	-	-

SAAQS: Saskatchewan Ambient Air Quality Standards




Exceedance Summary (24-Hour Average)									
24-hour Exceedance Information				Summary of other Parameters during Exceedance					
Pollutant	Date	Conc.		WS	WD	Precip.	SO ₂	H ₂ S	PM _{2.5}
H ₂ S	8/3/26	4.13		<Samp	<Samp	<Samp	0.70	4.13	<Samp
H ₂ S	8/8/26	4.22		<Samp	<Samp	<Samp	0.46	4.22	<Samp
H ₂ S	8/11/26	5.61		<Samp	<Samp	<Samp	1.27	5.61	<Samp


SESAA Wauchope Station - August 2026 Summary Report

Exceedance Summary (1-Hour Average)


1-hour Exceedance Information			Summary of Other Parameters During Exceedance					
Pollutant	Date and Time	Conc.	WS	WD	Precip.	SO ₂	H ₂ S	PM _{2.5}
H ₂ S	08/02/2026 24:00	14.63	NoData	NoData	NoData	0.6	14.63	NoData
H ₂ S	08/03/2026 05:00	11.95	NoData	NoData	NoData	0.8	11.95	NoData
H ₂ S	08/03/2026 07:00	11.24	NoData	NoData	NoData	0.8	11.24	NoData
H ₂ S	08/06/2026 02:00	14.7	NoData	NoData	NoData	0.7	14.7	NoData
H ₂ S	08/08/2026 07:00	17.48	NoData	NoData	NoData	0	17.48	NoData
H ₂ S	08/11/2026 03:00	12.11	NoData	NoData	NoData	0.6	12.11	NoData
H ₂ S	08/11/2026 04:00	15.15	NoData	NoData	NoData	0.8	15.15	NoData
H ₂ S	08/11/2026 05:00	16.15	NoData	NoData	NoData	0.7	16.15	NoData
H ₂ S	08/11/2026 06:00	17.9	NoData	NoData	NoData	0.9	17.9	NoData
H ₂ S	08/11/2026 07:00	14.84	NoData	NoData	NoData	0.8	14.84	NoData




Wind Rose



Pollution Rose SO₂



Pollution Rose H₂S



Pollution Rose PM_{2.5}

SESAA Wauchope Station - August 2026 Summary Report

Parameters	Unit	Color Code for Wind Speed and Pollutant Concentration Class					
Wind Speed	m/s	N/A	N/A	N/A	N/A	N/A	N/A
SO ₂	ppb	0.0-C-1.2	1.2-C-2.4	2.4-C-3.6	3.6-C-4.8	4.8-C-6.0	C > 6.0
H ₂ S	ppb	0.0-C-1.2	1.2-C-2.4	2.4-C-3.6	3.6-C-4.8	4.8-C-6.0	C > 6.0
PM _{2.5}	µg/m ³	N/A	N/A	N/A	N/A	N/A	N/A

Monthly Update


- * All data has been zero baseline corrected, when required. Data may change after final validation process.
- * All compliance parameters were within the Saskatchewan Ambient Air Quality Standards with the exception of three, 24-hour and ten 1-hour H₂S exceedances.
- * All compliance parameters were less than 90% operational for the month with the exception of PM_{2.5}.
- * Meteorological sensor removed for repair.
- * The PM_{2.5} analyzer has been malfunctioning since July 23, 2026, and no valid data are available. The instrument will be inspected during the next quarterly site visit.
- * One hour of data were flagged for all gas analyzers due to daily IZS checks, resulting in the loss of one hour of data each day.
- * The H₂S analyzer is currently under investigation; therefore, the exceedances have been flagged and are provided for reference only.
- * Pollution Rose data developed using wind data from the nearby Glen Ewan Station.











SESAA Weyburn Station - August 2026 Summary Report

Pollutants		Month Records		24 Hour Records				1 Hour Records			
Name	Unit	Average Concentration	Uptime	Maximum		SAAQS*	Exceedance Days	Maximum		SAAQS*	Exceed Hours
				Conc.	Date			Conc.	Date and Time		
NO	ppb	0.3	95.8%	0.8	8/18/2026	-	-	6.1	08/18/2026 07:00	-	-
NO ₂	ppb	0.3	95.8%	1.1	8/28/2026	106	0	6.4	08/20/2026 22:00	159	0
NO _x	ppb	0.5	95.8%	1.4	8/28/2026	-	-	8.3	08/18/2026 07:00	-	-
SO ₂	ppb	2.1	95.8%	4.3	8/27/2026	48	0	16.4	08/23/2026 08:00	172	0
H ₂ S	ppb	2.1	95.8%	3.4	8/27/2026	3.6	0	7.3	08/27/2026 03:00	11	0
PM _{2.5}	µg/m ³	No Data	0.0%	NoData	-	28	0	No Data	No Data	-	-

SAAQS: Saskatchewan Ambient Air Quality Standards



*If AQI cannot be calculated due to missing parameters, AQHI plus is reported using PM_{2.5}




SESAA Weyburn Station - August 2026 Summary Report

Exceedance Summary (24-Hours Average)

24-hours Exceedance Information			Summary of other Parameters during Exceedance							
Pollutant	Date	Conc.	WS	WD	H ₂ S	NO	NO ₂	NO _x	SO ₂	PM _{2.5}

Exceedance Summary (1-Hour Average)

1-hour Exceedance Information			Summary of Other Parameters During Exceedance					
Pollutant	Date and Time	Conc.	WS	WD	Precip.	SO ₂	NO ₂	PM _{2.5}



Parameters	Unit	Color Code for Wind Speed and Pollutant Concentration Class						
Wind Speed	m/s	0.5-W-1.2	1.2-W-2.4	2.4-W-3.6	3.6-W-4.8	4.8-W-6.0	W > 6.0	
NO ₂	ppb	0.0-C-1.2	1.2-C-2.4	2.4-C-3.6	3.6-C-4.8	4.8-C-6.0	C > 6.0	
SO ₂	ppb	0.0-C-0.4	0.4-C-0.8	0.8-C-1.2	1.2-C-1.6	1.6-C-2.0	C > 2.0	
H ₂ S	ppb	0.0-C-0.4	0.4-C-0.8	0.8-C-1.2	1.2-C-1.6	1.6-C-2.0	C > 2.0	
PM _{2.5}	µg/m ³	N/A	N/A	N/A	N/A	N/A	N/A	

Monthly Update

- * All data has been zero drift and baseline corrected, where required. Data may change after final validation process.
- * All compliance parameters were within the Saskatchewan Ambient Air Quality Standards.
- * All compliance parameters were more than 90% operational for the month with the exception of PM_{2.5}.
- * A total of 744 hours of PM_{2.5} data was flagged as "analyzer error," resulting in missing and flatlined data.
- * One hour of data were flagged for all gas analyzers due to daily IZS checks, resulting in the loss of one hour of data each day.
- * Most recent quarterly calibrations were completed on July 19, 2026.

